

Declaration of Conformity

Manufacturer: KD Scientific
Address: 84 October Hill Road, Holliston, Massachusetts 01746, USA

We herewith declare that the following products:

Product Name: Syringe Pump
Models: KD 100 Series KD 310 Series
78-9100 78-9311
78-9101
78-9120

Are in conformity of the following applicable European regulations and directives:

2014/35/EU Low Voltage directive (LVD)
2014/30/EU Electromagnetic Compatibility directive (EMC)
2012/19/EU Waste electrical and electronic equipment directive (WEEE)¹
2011/65/EU Restriction on the use of certain hazardous substances directive (RoHS)¹

Standards used to demonstrate conformity include:

EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use –EMC requirements²
EN 50581:2012 Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances¹
EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control and laboratory use, General Requirements

Conforms to UL61010-1:2012
Certified to CSA 22.2 No. 61010-1:2012



Report Numbers: 101450664BOX-003
3086793BOX-002

¹Classified as belonging to equipment categories 8 or 9

²Classified and tested as class A equipment in accordance with CISPR 11 definition in a basic electromagnetic environment. This equipment has also been tested and found to comply with the limits for a class A digital device, pursuant to CFR Title 47 part 15 of the FCC rules.

I, the undersigned, hereby declare that the equipment specified is in conformity with the relevant harmonised Union legislation. Signed for and on behalf of Harvard Apparatus at Holliston, USA

The logo features the lowercase letters 'kd' in a bold, blue, sans-serif font, followed by the word 'Scientific' in a larger, blue, serif font.

Signed:

A handwritten signature in black ink, appearing to read 'Mark Davis', with a horizontal line extending to the right.

Date:
September 29, 2017

Mark Davis
Director of Engineering