

# Declaration of Conformity

**Manufacturer:** KD Scientific, Inc.  
84 October Hill Road  
Holliston, Massachusetts 01746-1388, U.S.A.  
Phone: (508) 429-6809

We herewith declare that the following product:

**Product Name:** KDS Allegro Peristaltic Pumps  
**Model No.:** KDS Allegro Peristaltic Pump Series – Catalog No. 78-55xx

To which this declaration relates, is in conformity with the applicable EC Directives, harmonized standards, and other normative requirements:

**Application of Council Directive(s):**

2006/95/EC	Low Voltage Directive
2004/108/EC	Electromagnetic Compatibility Directive
2011/65/EU	RoHS Directive
2012/19/EU	WEEE Directive

**Standard(s) to which conformity is declared:**

**Safety:** IEC 61010-1:2010 (3rd Edition)  
UL 61010-1:2012 (3rd Edition)  
CSA C22.2# 61010-1:2012 (3rd Edition)

**Emissions/Immunity:** IEC 61326-1:2013  
FCC 47CFR 15B:2014, Class B

EMC and Safety compliance were evaluated by Intertek/ETL Semko

Reference test reports file numbers: 100898587 Box-002  
101276807 Box-003  
100179785 Box -005a, -005b



I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).

Date: July 20, 2015

Signed:

A handwritten signature in black ink, appearing to read 'M. ...'.

Harvard Bioscience Global Director  
Manufacturing and Process Engineering

The logo for kdScientific, with 'kd' in a bold, italicized blue font and 'Scientific' in a blue serif font.