

Declaration of Conformity

Manufacturer: KD Scientific
Address: 84 October Hill Road, Holliston, Massachusetts 01746-1388, USA

We herewith declare that the following products:

Product Name: KDS Allegro Peristaltic Pumps
Models: KDS Allegro Peristaltic Pump Series (78-55XX)

Are in conformity of the following applicable European regulations and directives:

| | |
|------------|--|
| 2014/35/EU | Low Voltage directive (LVD) |
| 2014/30/EU | Electromagnetic Compatibility directive (EMC) |
| 2012/19/EU | Waste electrical and electronic equipment directive (WEEE) ¹ |
| 2011/65/EU | Restriction on the use of certain hazardous substances directive (RoHS) ¹ |

Standards used to demonstrate conformity include:

| | |
|-----------------|--|
| EN 61326-1:2013 | Electrical equipment for measurement, control and laboratory use –EMC requirements ² |
| EN 61010-1:2010 | Safety requirements for electrical equipment for measurement, control and laboratory use, General Requirements |
| EN 50581:2012 | Assessment of products with respect to RoHS |

¹Classified as belonging to equipment categories 8 or 9

²Classified and tested as class A equipment in accordance with CISPR 11 definition in a basic electromagnetic environment. This equipment has also been tested and found to comply with the limits for a class A digital device, pursuant to CFR Title 47 part 15 of the FCC rules.

EMC and Safety compliance were evaluated by Intertek/ETL Semko

Reference test reports file numbers:

| |
|-----------------------------------|
| <i>100898587 Box-002</i> |
| <i>101276807 Box-003</i> |
| <i>100179785 Box -005a, -005b</i> |



I, the undersigned, hereby declare that the equipment specified is in conformity with the relevant harmonised Union legislation. Signed for and on behalf of Harvard Apparatus at Holliston, USA

The logo features the lowercase letters 'kd' in a bold, blue, sans-serif font, followed by the word 'Scientific' in a larger, blue, serif font.

Signed:

A handwritten signature in black ink, appearing to read 'Mark Davis'.

Date:

October 09, 2017

Mark Davis
Director of Engineering