## **Declaration of Conformity**

Manufacturer:	KD Scientific
Address:	84 October Hill Road, Holliston, Massachusetts 01746-1388, USA

We herewith declare that the following products:Product Name:KDS Legato 200 Series Syringe PumpsModels:KDS Legato 200 Series (78-82XX)

## Are in conformity of the following applicable European regulations and directives:

2014/35/EU	Low Voltage directive (LVD)
2014/30/EU	Electromagnetic Compatibility directive (EMC)
2012/19/EU	Waste electrical and electronic equipment directive (WEEE) <sup>1</sup>
2011/65/EU	Restriction on the use of certain hazardous substances directive (RoHs) $^{ m 1}$

## Standards used to demonstrate conformity include:

EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use –EMC requirements <sup>2</sup>
EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use, General
	Requirements
EN 50581:2012	Assessment of products with respect to RoHS

<sup>1</sup>Classified as belonging to equipment categories 8 or 9

<sup>2</sup>Classified and tested as class A equipment in accordance with CISPR 11 definition in a basic electromagnetic environment. This equipment has also been tested and found to comply with the limits for a class A digital device, pursuant to CFR Title 47 part 15 of the FCC rules.

EMC and Safety compliance were evaluated by Intertek/ETL Semko

Reference test reports file numbers:

100602834 Box-004, -004a 3189291 Box –001, -003, -005b



I, the undersigned, hereby declare that the equipment specified is in conformity with the relevant harmonised Union legislation. Signed for and on behalf of Harvard Apparatus at Holliston, USA



Signed:

Date: October 09, 2017

Mark Davis Director of Engineering

5617-002DC-Rev 1.1 KDS Legato 200 Pump Series DoC.docx