

Declaration of Conformity

Manufacturer: KD Scientific, Inc.
84 October Hill Road
Holliston, Massachusetts 01746-1388, U.S.A.
Phone: (508) 429-6809

We herewith declare that the following product:

Product Name: KDS Legato 100 Series Syringe Pump Series
Model No.: KDS Legato 100 Series - Catalog # 7881xx
KDS Legato 950 series- Catalog # 7889xx

To which this declaration relates, is in conformity with the applicable EC Directives, harmonized standards, and other normative requirements:

Application of Council Directive(s):

2006/95/EC	Low Voltage Directive
2004/108/EC	Electromagnetic Compatibility Directive
2011/65/EU	RoHS Directive
2012/19/EU	WEEE Directive

Standard(s) to which conformity is declared:

Safety: IEC 61010-1:2010 (3rd Edition)
UL 61010-1:2012 (3rd Edition)
CSA C22.2# 61010-1:2012 (3rd Edition)

Emissions/Immunity: IEC 61326-1:2013
FCC 47CFR 15B:2014, Class B

EMC and Safety compliance were evaluated by Intertek/ETL Semko

Reference test reports file numbers: 100602821 Box-001, -001a
100179785 Box -001, -003
100179785 Box -005a, -005b



I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).

Date: July 10, 2015

Signed:

A handwritten signature in black ink, appearing to read 'M. D. ...'.

Harvard Bioscience Global Director
Manufacturing and Process Engineering

The logo for kdScientific, with 'kd' in a bold, blue, sans-serif font and 'Scientific' in a blue, serif font.